



生物智造，助力卓越

Each molecule promises accuracy

检验报告

Certificate of Analysis

记录编号: R-QC-SMP-0013-03 版本号: 00

品名 Product Name	糖化血红蛋白(HbA1c)校准品 1 HbA1c Calibrator level 1		
批号 Batch Number	HH20220902	货号 Catalog Number	HH-CAL1001
生产日期 Production Date	2022.09.12	有效期至 Expiry Date	2023.09.11
储存条件 Storage Condition	2-8℃		
检验项目 Test Items	标准 Specification	结果 Result	
外观 Appearance	红棕色无定型粉末 Reddish brown amorphous powder	合格 PASS	
瓶间差 Difference between bottles	≤5% No more than 5%	合格 PASS	
含水量 Moisture content	≤6% No more than 6%	合格 PASS	
2-8℃稳定性 Stability at 2-8 °C	本产品复溶后在 2-8℃下储存 1 天，其检测值的相对偏差不超过 ± 10% Stable for 1 day at 2-8 °C after re-dissolution, and the relative deviation of the result comparing to initial value is no more than ± 10%		合格 PASS
冻融稳定性 Freeze-thaw stability	本产品复溶后再冻融 1 次，其检测值的相对偏差不超过 ± 10% If being frozen and thawed once again after re-dissolution, the relative deviation of the result comparing to initial value is no more than ± 10%		合格 PASS

结论:

Conclusion:



检验人/日期: 徐嘉帅 2022.09.26
Analyzed by/Date:

复核人/日期: 谭林 2022.09.26
Reviewed by/Date:

批准人/日期: 张俊 2022.09.26
Approved by/Date: